

(IPNXXXXXX)



1032 inch dia. spring-wire guide

Arrowg+ard Blue® One-Lumen CVC Kit



Contents:

- 1: Arrowg+ard Blue® One-Lumen Catheter: 16 Ga. (1.8 mm OD) x 16 cm
- Spring-Wire Guide, Marked: .032" (0.81 mm) dia. x 23-5/8" (60 cm) (Straight Soft Tip on One End - "J" Tip on Other) with Arrow Advancer
- Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW
- Introducer Needle
- Pressure Transduction Probe
- Introducer Needle: Echogenic 18 Ga. x 2-1/2" (6.35 cm) XTW and 5 mL Arrow® Raulerson Spring-Wire Introduction Syringe
- Injection Needle: SafetyGlide™1 25 Ga. x 1" (2.54 cm) and 3 mL Luer-
- Injection Needle: SafetyGlide™1 23 Ga. x 1-1/2" (3.81 cm) and 5 mL Luer-Slip Syringe
- Syringe: 10 mL Luer-Lock
- Tissue Dilator: 7 Fr. (2.3 mm) x 10.2 cm
- SecondSite™ Adjustable Hub: Fastener
- SecondSite™ Adjustable Hub: Catheter Clamp
- SharpsAway® II Locking Disposal Cup
- SharpsAway® Disposal Cup
- Maximal Barrier Drape™ with 4" fenestration

- Safety Scalpel: #11
- Checklist/CLIP Sheet
- Sterile Procedure Sign
- Gauze Pad: 2" x 2" (5 cm x 5 cm)


- Gauze Pad: 4" x 4" (10 cm x 10 cm)
- Surgical Apparel: Impervious Gown
- Dressing: Tegaderm®2 CHG 3-1/2" x 4-1/2" (8.5 cm x 11.5 cm)
- Surgical Apparel: Bouffant Cap
- Surgical Apparel: Mask with Eye Shield
- Suture: 3-0 Silk with Curved Needle
- Valve: MaxZero™3 Needleless Connector
- 1: HemoHopper® Fluid Receptacle
- ¹A trademark of Becton, Dickinson and Company.
- ²A registered trademark of 3M Company.
- ³A trademark of CareFusion or one of its subsidiaries

This catheter contains **CHLORHEXIDINE** and **SILVER SULFADIAZINE**



Gravity Priming Flow Volume Ratet (mL) (mL/hr)

Distal (16 Ga.) 2947



Store below 25°C (77°F). Avoid excessive heat above 40°C (104°F).

Contraindications: The Arrowg+ard Blue antimicrobial catheter is contraindicated for patients with known hypersensitivity to chlorhexidine. silver sulfadiazine and/or sulfa drugs.





LBL051404 R02 (2024-04)

^{*} Priming volumes are approximate and are done without accessories.

[†] Flow rate values are approximate and are determined using deionized water at 100 cm head