



REF AK-25123-F

12^{Fr.} 3^{Lumen}



Arrowg+ard Blue[®] Large-Bore Three-Lumen CVC for High Volume Infusions

Contents:

- Arrowg+ard Blue® Three-Lumen Catheter: 12 Fr. (4.1 mm OD) x 20 cm Spring-Wire Guide, Marked: .035" (0.89 mm) dia. x 26-3/4" (68 cm) (Straight Soft Tip on One End "J" Tip on Other) with Arrow Advancer
- Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer Needle 1:
- Injection Needle: 25 Ga. x 1" (2.54 cm) Blunt Fill Needle: 18 Ga. x 1-1/2" (3.81 cm)
- 1: 1: Pressure Transduction Probe
- 1: Introducer Needle: 18 Ga. x 2-1/2" (6.35 cm) XTW and 5 mL Arrow® Raulerson Spring-Wire Introduction
- Syringe Injection Needle: 22 Ga. x 1-1/2" (3.81 cm) and 5 mL Luer-Slip Syringe 1:
- Syringe: 3 mL Luer-Lock Tissue Dilator: 12 Fr. (4.0 mm) x 14 cm 5 mL 1% Lidocaine HCl Solution and Alcohol Prep 1:
- 3 mL Applicator 2% CHG and 70% IPA ChloraPrep^{®1} One-Step Solution with Hi-Lite Orange™ Tint SharpsAway[®] Disposal Cup Drape: 24" x 36" (61 cm x 91 cm) with 4" (10 cm) fenestration
- 1: Drape: 24" x 36" (61 cm x 91 cm) with a first Scalpel: #11 Gauze Pad: 2" x 2" (5 cm x 5 cm) Gauze Pad: 4" x 4" (10 cm x 10 cm) Suttre: 3-0 Braided Silk with Straight Needle
- 2
- 5:
- 1:
- ¹A registered trademark of CareFusion or one of its subsidiaries.

Rx only

- Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death.
- www.P65Warnings.ca.gov

Not made with natural rubber latex.

Store between 20 - 25°C (68 - 77°F)

Sterilized using ethylene oxide unless otherwise indicated in the contents list.

Fluid path components are non-pyrogenic

Contraindications: The Arrowg+ard Blue antimicrobial catheter is contraindicated for patients with known hypersensitivity to chlorhexidine, silver sulfadiazine and/or sulfa drugs.

Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	
Distal (16 Ga.)	0.40	2457	
Medial (12 Ga.)	1.3	10151	
Proximal (12 Ga.)	1.2	11039	

Priming volumes are approximate and are done without accessories.

† Flow rate values are approximate and are determined using deionized water at 100 cm head height.

LOT

 \sum








LBL052064 R01 (2022-02)





12^{Fr.} 3^{Lumen}





