



REF AK-04300

16 Ga. 1 Lumen

20 cm catheter length

.032 inch dia. spring-wire guide

# One-Lumen CVC

## Contents:

- 1: One-Lumen Catheter: 16 Ga. (1.8 mm OD) x 20 cm
- 1: Spring-Wire Guide: .032" (0.81 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End - "J" Tip on Other)
- 1: Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer Needle
- 1: Injection Needle: 25 Ga. x 1" (2.54 cm)
- 1: Blunt Fill Needle: 18 Ga. x 1-1/2" (3.81 cm)
- 1: Introducer Needle: 18 Ga. x 2-1/2" (6.35 cm) XTW and 5 mL Luer-Slip Syringe
- 1: Injection Needle: 20 Ga. x 1-1/2" (3.81 cm) and 5 mL Luer-Slip Syringe
- 1: Syringe: 3 mL Luer-Lock
- 1: Tissue Dilator: 7 Fr. (2.3 mm) x 10.2 cm
- 1: 5 mL 1% Lidocaine HCl Solution and Alcohol Prep
- 1: 3 mL Applicator 2% CHG and 70% IPA ChloraPrep®<sup>1</sup> One-Step Solution with Hi-Lite Orange™ Tint
- 1: SecondSite™ Adjustable Hub: Fastener
- 1: SecondSite™ Adjustable Hub: Catheter Clamp
- 1: SharpsAway® Disposal Cup
- 3: Drape: 26" x 18" (66 cm x 46 cm) with adhesive

- 1: Scalpel: #11
  - 3: Gauze Pad: 4" x 4" (10 cm x 10 cm)
  - 1: Suture: 3-0 Nylon with Straight Needle
  - 1: Tubing±: 6" Extension
- <sup>1</sup>A registered trademark of CareFusion or one of its subsidiaries.

## Rx only

**Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death.**

California Prop. 65

⚠ **WARNING:** Cancer and Reproductive Harm.  
www.P65Warnings.ca.gov

±**Precaution: Contains Phthalate: DEHP.**

**Warning: The results of certain animal experiments have shown phthalates to be potentially toxic to reproduction. Proceeding from the present state of scientific knowledge, risks for male premature infants cannot be excluded in the case of long-term exposure or application. Medical products containing phthalates should be used only temporarily with pregnant women, nursing mothers, babies and infants.**

**Not made with natural rubber latex.**

Store between 20 – 25°C (68 – 77°F).

Sterilized using ethylene oxide unless otherwise indicated in the contents list.

Fluid path components are non-pyrogenic.



Lumen

Priming Volume<sup>\*</sup> (mL)

Gravity Flow Rate<sup>†</sup> (mL/hr)

Distal (16 Ga.)

0.2

3179

<sup>\*</sup> Priming volumes are approximate and are done without accessories.

<sup>†</sup> Flow rate values are approximate and are determined using deionized water at 100 cm head height.



LOT



Arrow International LLC  
Subsidiary of Teleflex Incorporated  
3015 Carrington Mill Blvd.  
Morrisville, NC 27560 USA

**Teleflex®**  
www.teleflex.com

LBL053204 R00 (2021-11)



REF AK-04300

16 Ga. 1 Lumen

20 cm catheter length

.032 inch dia. spring-wire guide

# One-Lumen CVC