



REF **ASK-12703-SPH1** (IPNXXXXXX)

**7** Fr. **3** Lumen

**16** cm catheter length

**.032** inch dia. spring-wire guide



# Pressure Injectable Three-Lumen CVC Kit



**Contents:**

- 1: Three-Lumen Catheter: 7 Fr. (2.5 mm OD) x 16 cm, Pressure Injectable
- 1: Spring-Wire Guide, Marked: .032" (0.81 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End - "J" Tip on Other) with Arrow Advancer
- 1: Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer Needle
- 1: Pressure Transduction Probe
- 1: Introducer Needle: Echogenic 18 Ga. x 2-1/2" (6.35 cm) XTW and 5 mL Arrow® Raulerson Spring-Wire Introduction Syringe
- 1: Injection Needle: Eclipse™<sup>1</sup> 23 Ga. x 1-1/2" (3.81 cm) and 5 mL Luer-Slip Syringe
- 1: Injection Needle: Eclipse™<sup>1</sup> 25 Ga. x 1" (2.54 cm) and 3 mL Luer-Lock Syringe
- 1: Syringe: 10 mL Luer-Lock
- 1: Tissue Dilator: 8.5 Fr. (2.8 mm) x 10.2 cm
- 1: SecondSite™ Adjustable Hub: Fastener
- 1: SecondSite™ Adjustable Hub: Catheter Clamp
- 1: SharpsAway® II Locking Disposal Cup
- 1: SharpsAway® Disposal Cup
- 1: Maximal Barrier Drape™ with 4" fenestration
- 1: Towel
- 1: Needle Holder
- 1: Safety Scalpel: #11
- 1: Checklist/CLIP Sheet
- 1: Sterile Procedure Sign
- 2: Gauze Pad: 3" x 3" (7.6 cm x 7.6 cm)
- 2: Gauze Pad: 2" x 2" (5 cm x 5 cm)
- 5: Gauze Pad: 4" x 4" (10 cm x 10 cm)
- 1: Surgical Apparel: Impervious Gown
- 1: Dressing: I.V. Advanced 8.5 cm x 11.5 cm
- 1: Surgical Apparel: Bouffant Cap
- 1: Surgical Apparel: Mask with Eye Shield
- 1: Suture: 3-0 Silk with Curved Needle
- 3: Valve: MaxZero™<sup>2</sup> Needleless Connector
- 1: HemoHopper® Fluid Receptacle

<sup>1</sup>A trademark of Becton, Dickinson and Company.  
<sup>2</sup>A trademark of CareFusion or one of its subsidiaries.

Not made with natural rubber latex.



Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	MAX Pressure Injection Flow Rate** (mL/sec)
Distal (16 Ga.)	0.4	2947	10
Medial (18 Ga.)	0.4	1552	5
Proximal (18 Ga.)	0.4	1702	5

\* Priming volumes are approximate and are done without accessories.  
 † Flow rate values are approximate and are determined using deionized water at 100 cm head height.  
 \*\* Pressure injection flow rates are determined using media of 11.8 centipoise (cP) viscosity, with 152 cm administration set tubing.

