



REF PK-01552-WS

5 Fr. 2 Lumen 55 cm catheter length .018 inch dia. spring-wire guide

Two-Lumen PICC

Contents:

- 1: Two Lumen Catheter: 5 Fr. (1.76 mm OD) x 55 cm, T-Port Connector, Blue FlexTip®, Contamination Guard and Placement Wire
- 1: GlideThru™ Peel-Away Sheath: 5 Fr. x 2-3/4" (7 cm) Radiopaque over 5 Fr. Dilator
- 1: Spring-Wire Guide: .018" (0.46 mm) dia. x 13-1/8" (33 cm) (Straight Soft Tip on Both Ends)
- 1: Catheter: 22 Ga. x 1-3/4" (4.45 cm) Radiopaque over 25 Ga. RW Introducer Needle
- 1: Injection Needle: 18 Ga. x 1-1/2" (3.81 cm)
- 1: Introducer Needle: 21 Ga. x 1-1/2" (3.81 cm) TW
- 1: Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW
- 1: Syringe: 10 mL Luer-Slip
- 1: Skin Protectant Prep
- 1: 3 mL Applicator 2% CHG and 70% IPA Chloraprep®¹ One-Step Solution with Hi-Lite Orange™ Tint
- 2: Dust Cap: Non-Vented
- 1: SecondSite™ Adjustable Hub: Fastener
- 1: SecondSite™ Adjustable Hub: Catheter Clamp
- 1: SharpsAway® Disposal Cup
- 1: Towel
- 1: Towel: 24" x 36" (61 cm x 91 cm)
- 1: Drape: 24" x 36" (61 cm x 91 cm) with 4" (10 cm) fenestration
- 1: Forceps
- 1: Patient ID Card
- 1: Chart Sticker
- 1: Patient Information Booklet
- 1: Dressing: OpSite®² 14 cm x 10 cm
- 2: Paper Tape Measure
- 1: Tourniquet
- 5: Gauze Pad: 2" x 2" (5 cm x 5 cm)
- 6: Gauze Pad: 4" x 4" (10 cm x 10 cm)
- 1: Dressing: STATLOCK®³ Catheter Stabilization Device
- 2: Surgical Apparel: Mask
- 1: Tape: Steri-Strip®⁴

¹A registered trademark of CareFusion or one of its subsidiaries.
²A registered trademark of T.J. Smith & Nephew, Limited.
³A registered trademark of C. R. Bard, Inc.
⁴A registered trademark of 3M Company.

Rx only

Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. www.teleflex.com/IFU

California Prop. 65

⚠ WARNING: Cancer and Reproductive Harm. www.P65Warnings.ca.gov

Not made with natural rubber latex.

Store between 15 – 30°C (59 – 86°F). Avoid freezing and excessive heat above 40°C (104°F).

Fluid path components are non-pyrogenic.

Contraindications: The PICC is contraindicated wherever there is presence of device related infections or presence of thrombosis in the intended insertion vessel or catheter pathway. Clinical assessment of the patient must be completed to ensure no contraindications exist.



Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	Pump Flow Rate‡ (mL/hr)
Distal (18 Ga.)	0.4	380	2450
Proximal (20 Ga.)	0.4	106	1025

* Priming volumes are approximate and are done without accessories.
 † Flow rate values are approximate and are determined using deionized water at 100 cm head height.
 ‡ Pump flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.



LOT



Arrow International LLC
 Subsidiary of Teleflex Incorporated
 3015 Carrington Mill Blvd.
 Morrisville, NC 27560 USA

Teleflex®
www.teleflex.com

LBL056516 R01 (2022-02)



REF PK-01552-WS

5 Fr. 2 Lumen 55 cm catheter length .018 inch dia. spring-wire guide

Two-Lumen PICC