



(IPNXXXXXX)



55 cm catheter length

1018 inch dia. spring-wire guides

## **ARROW**

## Pressure Injectable Three-Lumen PICC with 130 cm Wire

- 1: Three Lumen Catheter: 6 Fr. (2.14 mm OD) x 55 cm, Pressure Injectable, Blue 1:

- Three Lumen Catheter: 6 Fr. (2.14 mm OD) x 55 cm, Pressure Injectable, Blue
  FlexTip®
  Peel-Away Sheath: 6.5 Fr. x 2-3/4" (7 cm) Radiopaque over 6.5 Fr. Dilator
  Spring-Wire Guide: .018" (0.46 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip
  On Both Ends)
  Spring-Wire Guide, Marked: .018" (0.46 mm) dia. x 51-3/16" (130 cm)(Straight
  Soft Tip on One End Straight Stiff Tip on Other)
  Introducer Needle: 21 Ga. x 1-1/2" (3.81 cm) TW
  Injection Needle: SafetyGlide™125 Ga. x 5/8" (1.60 cm)
  Injection Needle: SafetyGlide™121 Ga. x 1-1/2" (3.81 cm)
  Blunt Fill Needle: 18 Ga. x 1-1/2" (3.81 cm)
  Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW
  Syringe: 10 mL Luer-Lock
  Syringe: 3 mL Luer-Lock
  Syringe: 3 mL Luer-Lock
  Syringe: 3 mL Luer-Lock Pre-Filled Saline Syringe
  5 mL 1% Lidocaine HCl Solution and Alcohol Prep
  3 mL Applicator 2% CHG and 70% IPA ChloraPrep®2 One-Step Solution with
  Hi-Lite Orange™ Tint
  Dust Cap: Non-Vented
  SecondSite™ Adjustable Hub: Fastener

- SecondSite™ Adjustable Hub: Catheter Clamp SharpsAway® II Locking Disposal Cup SharpsAway® Disposal Cup Catheter Trimmer

- Catheter Trimmer
  Towel: 24" x 36" (61 cm x 91 cm)
  Safety Scalpel: #11
  Patient ID Card
  Chart Sticker
  Patient Information Booklet
  Medication Label: 1% Lidocaine
  Medication Label: 0.9% Saline
  Checklist/CLIP Sheet
  Paper Tape Measure
  Tourniquet

- 2. Paper Tape Weasure
  2. Tourniquet
  3. Gauze Pad: 2" x 2" (5 cm x 5 cm)
  4. Gauze Pad: 4" x 4" (10 cm x 10 cm)
  5. Dressing: I.V. Advanced 8.5 cm x 11.5 cm
  6. Dressing: BIOPATCH®3
  7. Dressing: STATLOCK®4 PICC Plus Catheter Stabilization Device
  7. Surgical Tape: Transpore™5 1" x 54" Roll

- <sup>1</sup>A trademark of Becton, Dickinson and Company.

  <sup>2</sup>A registered trademark of CareFusion or one of its subsidiaries.

  <sup>3</sup>A registered trademark of Johnson & Johnson Corporation.

  <sup>4</sup>A registered trademark of C. R. Bard, Inc.

  <sup>5</sup>A trademark of 3M Company.

Rx only
Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. www.teleflex.com/IFU
California Prop. 65

California Prop. 65

MARNING: Cancer and Reproductive Harm.
www.P65Warnings.ca.gov

Not made with natural rubber latex.

Store between 20 - 25°C (68 - 77°F)

Sterilized using ethylene oxide unless otherwise indicated in the contents

Fluid path components are non-pyrogenic

Arrow International provides the enclosed medication label(s) for your convenience. Please ensure that these labels are applied to the correct syringe and corresponding medication.

Contraindications: The Pressure Injectable PICC is contraindicated wherever there is presence of device related infections or presence of thrombosis in the intended insertion vessel or catheter pathway. Clinical assessment of the patient must be completed to ensure no contraindications exist

Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	Pump Flow Rate†† (mL/hr)	MAX Pressure Injection Flow Rate** (mL/sec)
Distal (17 Ga.)	0.68	996	5950	6
Medial (19 Ga.)	0.49	238	2540	not rated
Proximal (19 Ga.)	0.5	257	2500	not rated




















(01)XXXXXXXXXXXXXXX (17)YYMMDD (11)YYMMDD

Arrow International LLC Subsidiary of Teleflex Incorporated 3015 Carrington Mill Blvd. Morrisville, NC 27560 USA



I BI 056517 R05 (2024-03)





6 Fr. 3 Lumen 55 cm catheter length

1018 inch dia. spring-wire guides









<sup>\*</sup>Priming volumes are approximate and are done without accessories.

† Flow rate values are approximate and are determined using deionized water at 100 cm head height.

†† Pump flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.

\*\* Pressure injection flow rates are determined at the injector pressure setting of 300 psi maximum using media of 11.8 centipoise viscosity, with 152 cm pressure tubing.