



REF CDC-05531-PK1A (IPNXXXXXX)

3 Fr. 1 Lumen 55 cm catheter length .016 inch dia. spring-wire guide



One-Lumen PICC

Contents:

- 1: One Lumen Catheter: 3 Fr. (1.03 mm OD) x 55 cm, T-Port Connector, Contamination Guard and Placement Wire Origin: XX XXXX XX XXX XXXXXXXXXXXX
- 1: GlideThru™ Peel-Away Sheath: 3 Fr. x 2-3/4" (7 cm) Radiopaque over 3 Fr. Dilator Origin: XX XXXX XX XXX XXXXXXXXXXXX
- 1: Spring Wire Guide Assembly: .016" (0.40 mm) dia. x 13-1/8" (33 cm) (Straight Soft Tip On Both Ends) Origin: XX XXXX XX XXX XXXXXXXXXXXX
- 1: Introducer Needle: 21 Ga. x 1-1/2" (3.81 cm) TW Origin: XX XXXX XX XXX XXXXXXXXXXXX
- 1: Injection Needle: SafetyGlide™ 1 25 Ga. x 5/8" (1.60 cm) Origin: XX XXXX XX XXXXXXXXXXXX
- 1: Blunt Fill Needle: 18 Ga. x 1-1/2" (3.81 cm) Origin: XX XXXX XX XXX XXXXXXXXXXXX
- 1: Introducer Needle: Echoguard 21 Ga. x 2-3/4" (7 cm) TW Origin: XX XXXX XX XXXXXXXXXXXX
- 1: Syringe: 10 mL Luer-Lock Origin: XX XXXX XX XXX XXXXXXXXXXXX
- 1: Syringe: 3 mL Luer-Lock Origin: XX XXXX XX XXX XXXXXXXXXXXX
- 2: 10 mL Luer-Lock Pre-Filled Saline Syringe Origin: XX XXXX XX XXX XXXXXXXXXXXX
- 1: 5 mL 1% Lidocaine HCl Solution and Alcohol Prep Origin: XX XXXX XX XXX XXXXXXXXXXXX
- 1: 3 mL Applicator 2% CHG and 70% IPA Chloraprep®2 One-Step Solution with Hi-Lite Orange™ Tint Origin: XX XXXX XX XXX XXXXXXXXXXXX
- 1: Dust Cap: Non-Vented Origin: XX XXXX XX XXX XXXXXXXXXXXX
- 1: SharpsAway® II Locking Disposal Cup Origin: XX XXXX XX XXX XXXXXXXXXXXX
- 1: SharpsAway® Disposal Cup Origin: XX XXXX XX XXX XXXXXXXXXXXX
- 1: Catheter Trimmer Origin: XX XXXX XX XXX XXXXXXXXXXXX
- 1: Towel: 24" x 36" (61 cm x 91 cm) Origin: XX XXXX XX XXX XXXXXXXXXXXX
- 1: Maximal Barrier Drape™ with two 13 cm (5") x 15 cm (6") fenestrations Origin: XX XXXX XX XXX XXXXXXXXXXXX
- 1: Towel Origin: XX XXXX XX XXX XXXXXXXXXXXX
- 1: Safety Scalpel: #11 Origin: XX XXXX XX XXX XXXXXXXXXXXX
- 1: Sterile Procedure Sign Origin: XX XXXX XX XXX XXXXXXXXXXXX
- 1: Patient ID Card Origin: XX XXXX XX XXX XXXXXXXXXXXX
- 1: Chart Sticker Origin: XX XXXX XX XXX XXXXXXXXXXXX
- 1: Patient Information Booklet Origin: XX XXXX XX XXX XXXXXXXXXXXX
- 1: Medication Label: 1% Lidocaine Origin: XX XXXX XX XXX XXXXXXXXXXXX
- 1: Checklist/CLIP Sheet Origin: XX XXXX XX XXX XXXXXXXXXXXX
- 2: Paper Tape Measure Origin: XX XXXX XX XXX XXXXXXXXXXXX
- 1: Tourniquet Origin: XX XXXX XX XXX XXXXXXXXXXXX
- 5: Gauze Pad: 2" x 2" (5 cm x 5 cm) Origin: XX XXXX XX XXX XXXXXXXXXXXX
- 10: Gauze Pad: 4" x 4" (10 cm x 10 cm) Origin: XX XXXX XX XXX XXXXXXXXXXXX
- 1: Surgical Apparel: Impervious Gown Origin: XX XXXX XX XXX XXXXXXXXXXXX
- 1: Dressing: I.V. Advanced 8.5 cm x 11.5 cm Origin: XX XXXX XX XXX XXXXXXXXXXXX
- 1: Dressing: BIOPATCH®3 Origin: XX XXXX XX XXX XXXXXXXXXXXX
- 1: Surgical Apparel: Bouffant Cap Origin: XX XXXX XX XXX XXXXXXXXXXXX
- 1: Surgical Apparel: Mask with Eye Shield Origin: XX XXXX XX XXX XXXXXXXXXXXX
- 1: Dressing: STATLOCK®4 PICC Plus Catheter Stabilization Device Origin: XX XXXX XX XXX XXXXXXXXXXXX
- 1: Surgical Apparel: Mask Origin: XX XXXX XX XXX XXXXXXXXXXXX
- 1: Tape: Steri-Strip®5 Origin: XX XXXX XX XXX XXXXXXXXXXXX
- 1: HemoHopper® Fluid Receptacle Origin: XX XXXX XX XXX XXXXXXXXXXXX

¹A trademark of Becton, Dickinson and Company.
²A registered trademark of CareFusion or one of its subsidiaries.
³A registered trademark of Johnson & Johnson Corporation.
⁴A registered trademark of C. R. Bard, Inc.
⁵A registered trademark of 3M Company.



Lumen

	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	Pump Flow Rate‡ (mL/hr)
Distal (20 Ga.)	0.3	103	1044

* Priming volumes are approximate and are done without accessories.
 † Flow rate values are approximate and are determined using deionized water at 100 cm head height.
 ‡ Pump flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.

Not made with natural rubber latex.

Store between 20 – 25°C (68 – 77°F).
 Arrow International provides the enclosed medication label(s) for your convenience. Please ensure that these labels are applied to the correct syringe and corresponding medication.

Contraindications: The PICC is contraindicated wherever there is presence of device related infections or presence of thrombosis in the intended insertion vessel or catheter pathway. Clinical assessment of the patient must be completed to ensure no contraindications exist.

