PB-10000-850 8.5 X 10.5

REF CDC-05531-PK1A (IPNXXXXXX)

3 Fr. 1 Lumen 55 cm catheter length .016 inch dia. spring-wire guide

ARROW ErgoPack®

One-Lumen PICC

Contents:

1: One Lumen Catheter: 3 Fr. (1.03 mm OD) x 55 cm, T-Port Connector, Contamination Guard and Placement Wire
1: GlideThru™ Peel-Away Sheath: 3 Fr. x 2-3/4" (7 cm) Radiopaque over 3 Fr. Dilator
1: Spring Wire Guide Assembly: .016" (0.40 mm) dia. x 13-1/8" (33 cm) (Straight Soft Tip On Both Ends)
1: Introducer Needle: 21 Ga. x 1-1/2" (3.81 cm) TW
1: Injection Needle: SafetyGlide™[1] 25 Ga. x 5/8" (1.60 cm)
1: Blunt Fill Needle: 18 Ga. x 1-1/2" (3.81 cm)
1: Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW
1: Syringe: 10 mL Luer-Lock
1: Syringe: 3 mL Luer-Lock
2: 10 mL Luer-Lock Pre-Filled Saline Syringe
1: 5 mL 1% Lidocaine HCl Solution and Alcohol Prep
1: 3 mL Applicator 2% CHG and 70% IPA ChloraPrep®[2] One-Step Solution with Hi-Lite Orange™ Tint
1: Dust Cap: Non-Vented
1: SharpsAway® II Locking Disposal Cup
1: SharpsAway® Disposal Cup
1: Catheter Trimmer
1: Towel: 24" x 36" (61 cm x 91 cm)
1: Maximal Barrier Drape™ with two 13 cm (5") x 15 cm (6") fenestrations
1: Towel
1: Safety Scalpel: #11
1: Sterile Procedure Sign
1: Patient ID Card
1: Chart Sticker
1: Patient Information Booklet
1: Medication Label: 1% Lidocaine
1: Checklist/CLIP Sheet
2: Paper Tape Measure
1: Tourniquet
5: Gauze Pad: 2" x 2" (5 cm x 5 cm)
10: Gauze Pad: 4" x 4" (10 cm x 10 cm)
1: Surgical Apparel: Impervious Gown
1: Dressing: I.V. Advanced 8.5 cm x 11.5 cm
1: Dressing: BIOPATCH®[3]
1: Surgical Apparel: Bouffant Cap
1: Surgical Apparel: Mask with Eye Shield
1: Dressing: STATLOCK®[4] PICC Plus Catheter Stabilization Device
1: Surgical Apparel: Mask
1: Tape: Steri-Strip®[5]
1: HemoHopper® Fluid Receptacle

[1]A trademark of Becton, Dickinson and Company.
[2]A registered trademark of CareFusion or one of its subsidiaries.
[3]A registered trademark of Johnson & Johnson Corporation.
[4]A registered trademark of C. R. Bard, Inc.
[5]A registered trademark of 3M Company.

Not made with natural rubber latex.

Store between 20 – 25°C (68 – 77°F).

Arrow International provides the enclosed medication label(s) for your convenience. Please ensure that these labels are applied to the correct syringe and corresponding medication.

Contraindications: The PICC is contraindicated wherever there is presence of device related infections or presence of thrombosis in the intended insertion vessel or catheter pathway. Clinical assessment of the patient must be completed to ensure no contraindications exist.

Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	Pump Flow Rate†† (mL/hr)
Distal (20 Ga.)	0.3	103	1044

* Priming volumes are approximate and are done without accessories.
† Flow rate values are approximate and are determined using deionized water at 100 cm head height.
†† Pump flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.

LATEX 20°C (68°F) 25°C (77°F)

LBL056551 R04 (2024-02)