



REF PR-45563-HPHNL

6 Fr. 3 Lumen 55 cm catheter length .018 inch dia. spring-wire guide

# Pressure Injectable Arrowg+ard Blue Advance® Three-Lumen PICC with 130 cm Hydrophilic-Coated Nitinol Guidewire

## Contents:

- 1: Three Lumen TaperFree® Catheter with Arrowg+ard Blue Advance® Antimicrobial/Antithrombogenic Protection<sup>1</sup>: 6 Fr. (2.11 mm OD) x 55 cm, Pressure Injectable, Blue FlexTip®  
 1: GlideThru™ Peel-Away Sheath: 6 Fr. x 4" (10 cm) Radiopaque over 6 Fr. Dilator  
 1: Spring-Wire Guide, Nitinol, Marked: .018" (0.46 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End - Straight Stiff Tip on Other)  
 1: Guidewire, Marked: .018" (0.46 mm) dia. x 130 cm Hydrophilic-Coated Nitinol with Soft Tip Tungsten Coil  
 1: Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW  
 1: Syringe: 10 mL Luer-Lock  
 1: Skin Protectant Prep  
 3: Dust Cap: Non-Vented  
 1: SecondSite™ Adjustable Hub: Fastener  
 1: SecondSite™ Adjustable Hub: Catheter Clamp  
 1: Catheter Trimmer  
 1: Safety Scalpel: #11  
 1: Patient ID Card  
 1: Chart Sticker  
 1: Patient Information Booklet  
 1: Checklist/CLIP Sheet  
 1: Paper Tape Measure  
 1: Dressing: STATLOCK®<sup>2</sup> Catheter Stabilization Device

<sup>1</sup>Licensed under US Patent No. 7,329,412.

<sup>2</sup>A registered trademark of C. R. Bard, Inc.

## Rx only

**Warning:** Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. [www.teleflex.com/IFU](http://www.teleflex.com/IFU)

California Prop. 65

**⚠ WARNING:** Cancer and Reproductive Harm.  
[www.P65Warnings.ca.gov](http://www.P65Warnings.ca.gov)

**Not made with natural rubber latex.**

Store below 25°C (77°F). Avoid excessive heat above 30°C (86°F).

Fluid path components are non-pyrogenic.

**Contraindications:** The Pressure Injectable Arrowg+ard Blue Advance antimicrobial/antithrombogenic catheter is contraindicated:

- for patients with known hypersensitivity to chlorhexidine
- in the presence of device related infection in the intended insertion vessel or catheter pathway
- in the presence of thrombosis in the intended insertion vessel or catheter pathway



Lumen

	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	Pump Flow Rate†† (mL/hr)	MAX Pressure Injection Flow Rate** (mL/sec)
Distal (18 Ga.)	0.61	880	5360	6
Medial (19 Ga.)	0.45	244	2090	not rated
Proximal (19 Ga.)	0.47	257	2160	not rated

\* Priming volumes are approximate and are done without accessories.

† Flow rate values are approximate and are determined using deionized water at 100 cm head height.

†† Pump flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.

\*\* Pressure injection flow rates are determined at the injector pressure setting of 300 psi maximum using media of 11.8 centipoise viscosity, with 152 cm pressure tubing.



LOT



Arrow International LLC  
 Subsidiary of Teleflex Incorporated  
 3015 Carrington Mill Blvd.  
 Morrisville, NC 27560 USA

**Teleflex®**  
[www.teleflex.com](http://www.teleflex.com)

LBL057538 R00 (2021-10)