



REF SP-09801 (IPNXXXXXX)

8.5 Fr. 10 cm sheath length .035 inch dia. spring-wire guide

# PSI Kit for use with 7 - 7.5 Fr. Catheters

**Contents:**

- 1: Arrow-Flex® Sheath: 8.5 Fr. (2.8 mm tip ID) x 10 cm and Tissue Dilator origin:xxxxxxxxxxxxxxxxxxxx
- 1: Spring-Wire Guide: .035" (0.89 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End - "J" Tip on Other) origin:xxxxxxxxxxxxxxxxxxxx
- 1: Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer Needle origin:xxxxxxxxxxxxxxxxxxxx
- 1: Injection Needle: 25 Ga. x 1" (2.54 cm) origin:xxxxxxxxxxxxxxxxxxxx
- 1: Injection Needle: 22 Ga. x 1-1/2" (3.81 cm) origin:xxxxxxxxxxxxxxxxxxxx
- 1: Introducer Needle: 18 Ga. x 2-1/2" (6.35 cm) XTW origin:xxxxxxxxxxxxxxxxxxxx
- 1: Syringe: 3 mL Luer-Lock origin:xxxxxxxxxxxxxxxxxxxx
- 2: Syringe: 5 mL Luer-Slip origin:xxxxxxxxxxxxxxxxxxxx
- 1: Obturator: 8 Fr. origin:xxxxxxxxxxxxxxxxxxxx
- 1: Cath-Gard® Catheter Contamination Shield: 80 cm with Tuohy-Borst Adapter, Sterile Tape origin:xxxxxxxxxxxxxxxxxxxx
- 1: Dust Cap: Non-Vented origin:xxxxxxxxxxxxxxxxxxxx
- 1: SharpsAway® Disposal Cup origin:xxxxxxxxxxxxxxxxxxxx
- 1: Drape: 24" x 36" (61 cm x 91 cm) with 4" (10 cm) fenestration origin:xxxxxxxxxxxxxxxxxxxx
- 1: Scalpel: #11 origin:xxxxxxxxxxxxxxxxxxxx
- 1: Stopcock: 3-way High-Flow origin:xxxxxxxxxxxxxxxxxxxx
- 2: Prep Sponge Swab origin:xxxxxxxxxxxxxxxxxxxx
- 2: Gauze Pad: 2" x 2" (5 cm x 5 cm) origin:xxxxxxxxxxxxxxxxxxxx
- 5: Gauze Pad: 4" x 4" (10 cm x 10 cm) origin:xxxxxxxxxxxxxxxxxxxx
- 1: Suture: 3-0 Braided Silk with Straight Needle origin:xxxxxxxxxxxxxxxxxxxx
- 1: Tray: Prep origin:xxxxxxxxxxxxxxxxxxxx

**Rx only**

**Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death.**

**California Prop. 65**

**WARNING:** Cancer and Reproductive Harm.  
[www.P65Warnings.ca.gov](http://www.P65Warnings.ca.gov)

**Caution: This Product Contains Dry Natural Rubber.**

Sterilized using ethylene oxide unless otherwise indicated in the contents list.

Fluid path components are non-pyrogenic.



**LOT** Lot Number

YYYY-MM-DD

YYYY-MM-DD



(01)XXXXXXXXXXXX  
(17)YYMMDD  
(11)YYMMDD  
(10)Lot Number

Arrow International LLC  
Subsidiary of Teleflex Incorporated  
3015 Carrington Mill Blvd.  
Morrisville, NC 27560 USA

Set Product of xxx  
Packaged in xxx



LBL058343 R02 (2024-06)



REF SP-09801

8.5 Fr. 10 cm sheath length .035 inch dia. spring-wire guide

# PSI Kit for use with 7 - 7.5 Fr. Catheters

