







(IPNXXXXXX)

9 Fr. distal lumen

10 cm device length

.035 inch dia. spring-wire quide

## Arrowg+ard Blue® MAC™ Two-Lumen Central Venous Access for use with 7 - 8 Fr. **Catheters**

## Contents:

- origin:xxxxxxxxxxxxxxxxxxxxxx
- origin:xxxxxxxxxxxxxxxxxxxxxxx

- Licensed under US Patent No. 6,706,024
- <sup>2</sup>A registered trademark of CareFusion or one of its subsidiaries.

Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death.

injury or death.

California Prop. 65

WARNING: Cancer and Reproductive Harm.

www.P65Warnings.ca.gov

Precaution: Contains Phthalate: DEHP.

Warning: The results of certain animal experiments have shown phthalates to be potentially toxic to reproduction. Proceeding from the present state of scientific knowledge, risks for male premature infants cannot be excluded in the case of long-term exposure or application. Medical products containing phthalates should be used only temporarily with pregnant women, nursing mothers, babies and infants.

Not made with natural rubber latex.

Store between 20 – 25°C (68 – 77°F).

Sterilized using ethylene oxide unless otherwise indicated in the contents

Store between 20 - 25°C (68 - 77°F).

Contraindications: The Arrowg+ard Blue antimicrobial catheter is contraindicated for patients with known hypersensitivity to chlorhexidine acetate, silver sulfadiazine, and/or sulfa drugs.

Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)
Distal (9 Fr.) no catheter	1.65	29310
Distal (9 Fr.) with 8 Fr. catheter	0.9	6720
Provimal (12 Ca.) no cathotor	0.6	6710

<sup>\*</sup> Priming volumes are approximate and are done without accessories.



Arrow International LLC Subsidiary of Teleflex Incorporated 3015 Carrington Mill Blvd. Morrisville, NC 27560 USA


**Teleflex**®

LBL058359 R01 (2024-04)



**LOT** Lot Number

YYYY-MM-DD



YYYY-MM-DD







1035 inch dia. spring-wire





(01)XXXXXXXXXXXXXXX (17)YYMMDD (11)YYMMDD

(10)Lot Number



<sup>†</sup> Flow rate values are approximate and are determined using deionized water at 100 cm head height.