



REF **AK-11142-SPCS**

**9** Fr. distal lumen

**2** Lumen

**10** cm device length

**.035** inch dia. spring-wire guide

# MAC™ Two-Lumen Central Venous Access for use with 7 - 7.5 Fr. Catheters

**Contents:**

- |   |  |
|---|--|
| 1: Two-Lumen Central Venous Access Device: 4" (10 cm) Radiopaque Polyurethane with 9 Fr. Distal Lumen, Integral Hemostasis Valve, Extension Line Clamps | 1: Sterile Tape  |
| 1: Spring-Wire Guide, Marked: .035" (0.89 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End - "J" Tip on Other) with Arrow Advancer              | 1: Dust Cap: Vented  |
| 1: Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer Needle  | 2: Dust Cap: Non-Vented  |
| 1: Injection Needle: 25 Ga. x 1" (2.54 cm)  | 1: SharpsAway® II Locking Disposal Cup                           |
| 1: Blunt Fill Needle: 18 Ga. x 1-1/2" (3.81 cm)   | 1: SharpsAway® Disposal Cup                                      |
| 1: Pressure Transduction Probe  | 1: Drape: 24" x 36" (61 cm x 91 cm) with 4" (10 cm) fenestration |
| 1: Introducer Needle: 18 Ga. x 2-1/2" (6.35 cm) XTW and 5 mL Arrow® Raulerson Spring-Wire Introduction Syringe  | 1: Drape: 24" x 36" (61 cm x 91 cm)                              |
| 1: Injection Needle: 22 Ga. x 1-1/2" (3.81 cm) and 5 mL Luer-Slip Syringe   | 1: Safety Scalpel: #11   |
| 1: Syringe: 3 mL Luer-Lock  | 1: Stopcock: 4-way High-Flow                                     |
| 1: Tissue Dilator: 9 Fr. (3.0 mm) x 20.7 cm   | 2: Gauze Pad: 2" x 2" (5 cm x 5 cm)                              |
| 1: 5 mL 1% Lidocaine HCl Solution and Alcohol Prep  | 5: Gauze Pad: 4" x 4" (10 cm x 10 cm)                            |
| 1: 3 mL Applicator 2% CHG and 70% IPA Chloraprep® <sup>1</sup> One-Step Solution with Hi-Lite Orange™ Tint  | 1: Suture: 3-0 Braided Silk with Straight Needle                 |
| 1: Obturator: 8 Fr.   | 1: Tubing‡: 20" (50 cm) Extension                                |
| 1: Cath-Gard® Catheter Contamination Shield: 80 cm with Tuohy-Borst Adapter,  |  |

<sup>1</sup>A registered trademark of CareFusion or one of its subsidiaries.

**Rx only**

**Warning:** Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. [www.teleflex.com/IFU](http://www.teleflex.com/IFU)

California Prop. 65

**⚠ WARNING:** Cancer and Reproductive Harm. [www.P65Warnings.ca.gov](http://www.P65Warnings.ca.gov)

**±Precaution:** Contains Phthalate: DEHP.

**Warning:** The results of certain animal experiments have shown phthalates to be potentially toxic to reproduction. Proceeding from the present state of scientific knowledge, risks for male premature infants cannot be excluded in the case of long-term exposure or application. Medical products containing phthalates should be used only temporarily with pregnant women, nursing mothers, babies and infants.

Not made with natural rubber latex.

Store between 20 – 25°C (68 – 77°F).

Sterilized using ethylene oxide unless otherwise indicated in the contents list.

Fluid path components are non-pyrogenic.



Lumen

Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)
Distal (9 Fr.) no catheter	1.65	31440
Distal (9 Fr.) with 8 Fr. catheter	0.9	9040
Proximal (12 Ga.) no catheter	0.6	8970

\* Priming volumes are approximate and are done without accessories.

† Flow rate values are approximate and are determined using deionized water at 100 cm head height.



LOT



Arrow International LLC  
Subsidiary of Teleflex Incorporated  
3015 Carrington Mill Blvd.  
Morrisville, NC 27560 USA

**Teleflex®**  
[www.teleflex.com](http://www.teleflex.com)

LBL058364 R00 (2022-02)



REF **AK-11142-SPCS**

**9** Fr. distal lumen

**2** Lumen

**10** cm device length

**.035** inch dia. spring-wire guide

# MAC™ Two-Lumen Central Venous Access for use with 7 - 7.5 Fr. Catheters