ARROW

Product of xxx Packaged in xxx





REF CS-22122-F (IPNXXXXXXX)

16 cm catheter length .035 inch dia. spring-wire guide

Arrowg+ard Blue® Two-Lumen Hemodialysis Catheter for High **Volume Infusions**

- Two-Lumen Catheter: 12 Fr. (4.06 mm OD) x 16 cm Radiopaque Polyurethane with Blue
- FlexTip®, Arrowg+ard® Antimicrobial Surface Treatment, Extension Line Clamps
 Spring-Wire Guide, Marked: .035" (0.89 mm) dia. x 23-5/8" (60 cm) (Straight Soft Tip on One End "J" Tip on Other) with Arrow Advancer
 Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer Needle
 Processor Transduction Brobe
- Pressure Transduction Probe
 Introducer Needle: 18 Ga. x 2-1/2" (6.35 cm) XTW and 5 mL Arrow® Raulerson Spring-
- Wire Introduction Syringe
 Tissue Dilator: 12 Fr. (4.0 mm) x 14 cm
- Dust Cap: Non-Vented
 Drape: 24" x 36" (61 cm x 91 cm)
 Scalpel: #11

All components are CE 2797 unless otherwise noted.

Rx	on	ly
Ma	rni	n

ng: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. www.teleflex.com/IFU California Prop. 65

M WARNING: Cancer and Reproductive Harm.

www.P65Warnings.ca.gov
Not made with natural rubber latex.

Store below 25°C (77°F). Avoid excessive heat above 40°C (104°F). Fluid path components are non-pyrogenic.

Contraindications: The Arrow large-bore two-lumen catheter is not designed for long-term (2 30 days) hemodialysis or for use in patients with thrombosed vessels.

The Arrowg+ard Blue antimicrobial catheter is contraindicated for patients with known hypersensitivity to chlorhexidine, silver sulfadiazine, and/or sulfa drugs.


Lumen	Priming Volume* (mL)	Flow Rate (mL/min)	Venous Pressure Setting†† (mmHg)
Distal (12 Ga.)	1.4	220	100
Proximal (12 Ga.)	1.3	370 220	200 100
		370	200

^{*} Priming volumes are approximate and are done without accessories













(01)XXXXXXXXXXXXXXX (17)YYMMDD (11)YYMMDD (10)Lot Number



FLI Authorized Representative and Importer:



Teleflex Medical IDA Business and Technology Park Dublin Road, Athlone
Co. Westmeath, Ireland

Subsidiary of Teleflex Incorporated 3015 Carrington Mill Blvd. Morrisville, NC 27560 USA



LBL058884 R01 (2024-01) /









16 cm catheter length

.035 inch dia. spring-wire guide



^{††} Blood flow rates are approximate and were determined by varying the venous pressure setting from 100 to 200 mmHg in an in vitro experiment using blood analog solution.