



REF **SI-11242** (IPNXXXXXX)

**9** Fr. distal lumen    **2** Lumen    **11.5** cm device length    **.035** inch dia. spring-wire guide

# MAC™ Two-Lumen Central Venous Access Set with Integral Hemostasis Valve for use with 7 - 7.5 Fr. Catheters

**Contents:**

- 1: Two-Lumen Central Venous Access Device: 4-1/2" (11.5 cm) Radiopaque Polyurethane with 9 Fr. Distal Lumen, Integral Hemostasis Valve, Extension Line Clamps
- 1: Spring-Wire Guide, Marked: .035" (0.89 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End - "J" Tip on Other) with Arrow Advancer
- 1: Introducer Needle: 18 Ga. x 2-1/2" (6.35 cm) XTW
- 1: Pressure Transduction Probe
- 1: Arrow® Raulerson Spring-Wire Introduction Syringe: 5 mL
- 1: Tissue Dilator
- 1: Obturator: 8 Fr.
- 1: Cath-Gard® Catheter Contamination Shield: 80 cm with Tuohy-Borst Adapter, Sterile Tape
- 1: Dust Cap: Vented
- 3: Dust Cap: Non-Vented
- 1: Stopcock: 4-way High-Flow
- 3: Gauze Pad: 4" x 4" (10 cm x 10 cm)

All components are CE 2797 unless otherwise noted.

**Rx only**

Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. [www.teleflex.com/IFU](http://www.teleflex.com/IFU)

California Prop. 65

⚠ **WARNING:** Cancer and Reproductive Harm.

[www.P65Warnings.ca.gov](http://www.P65Warnings.ca.gov)

Not made with natural rubber latex.

Fluid path components are non-pyrogenic.



Lumen

Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)
Distal (9 Fr.) no catheter	1.8	31440
Distal (9 Fr.) with 7 Fr. catheter	1.45	16280
Distal (9 Fr.) with 8 Fr. catheter	1.15	9040
Proximal (12 Ga.) no catheter	0.6	8970

\* Priming volumes are approximate and are done without accessories.

† Flow rate values are approximate and are determined using deionized water at 100 cm head height.

en cs da nl et fi fr de el hu it no pl pt ru sk sl es  
sv tr



CAS # 7440-48-4

LOT Lot Number

YYYY-MM-DD

YYYY-MM-DD



(01)XXXXXXXXXXXXXX  
(17)YYMMDD  
(11)YYMMDD  
(10)Lot Number

Product of xxx  
Packaged in xxx



EC REP Teleflex Medical  
IDA Business and  
Technology Park  
Dublin Road, Athlone  
2797 Co. Westmeath, Ireland

All components are CE 2797 unless otherwise noted.

Arrow International LLC  
Subsidiary of Teleflex Incorporated  
3015 Carrington Mill Blvd.  
Morrisville, NC 27560 USA

**Teleflex**  
www.teleflex.com

LBL060779 R02 (2025-02)



# MAC™ Two-Lumen Central Venous Access Set with Integral Hemostasis Valve for use with 7 - 7.5 Fr. Catheters



REF

SI-11242

**9** Fr. distal lumen    **2** Lumen    **11.5** cm device length    **.035** inch dia. spring-wire guide