

Product of xxx  
Packaged in xxx



REF **SI-11242** (IPNXXXXXX)

**9** Fr. distal lumen    **2** Lumen    **11.5** cm device length    **.035** inch dia. spring-wire guide

# MAC™ Two-Lumen Central Venous Access Set with Integral Hemostasis Valve for use with 7 - 7.5 Fr. Catheters

**Contents:**

- 1: Two-Lumen Central Venous Access Device: 4-1/2" (11.5 cm) Radiopaque Polyurethane with 9 Fr. Distal Lumen, Integral Hemostasis Valve, Extension Line Clamps
- 1: Spring-Wire Guide, Marked: .035" (0.89 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End - "J" Tip on Other) with Arrow Advancer
- 1: Introducer Needle: 18 Ga. x 2-1/2" (6.35 cm) XTW
- 1: Pressure Transduction Probe
- 1: Arrow® Raulerson Spring-Wire Introduction Syringe: 5 mL
- 1: Tissue Dilator
- 1: Obturator: 8 Fr.
- 1: Cath-Gard® Catheter Contamination Shield: 80 cm with Tuohy-Borst Adapter, Sterile Tape
- 1: Dust Cap: Vented
- 3: Dust Cap: Non-Vented
- 1: Stopcock: 4-way High-Flow
- 3: Gauze Pad: 4" x 4" (10 cm x 10 cm)

All components are CE 2797 unless otherwise noted.

**Rx only**

Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. [www.teleflex.com/IFU](http://www.teleflex.com/IFU)

California Prop. 65

⚠ **WARNING:** Cancer and Reproductive Harm.

[www.P65Warnings.ca.gov](http://www.P65Warnings.ca.gov)

Not made with natural rubber latex.

Fluid path components are non-pyrogenic.



Lumen

	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)
Distal (9 Fr.) no catheter	1.8	31440
Distal (9 Fr.) with 7 Fr. catheter	1.45	16280
Distal (9 Fr.) with 8 Fr. catheter	1.15	9040
Proximal (12 Ga.) no catheter	0.6	8970

\* Priming volumes are approximate and are done without accessories.

† Flow rate values are approximate and are determined using deionized water at 100 cm head height.



LOT LotNumber  
YYYY-MM-DD    YYYY-MM-DD



(01)XXXXXXXXXXXXXX  
(17)YYMMDD  
(11)YYMMDD  
(10)LotNumber

EU Authorized Representative and Importer:  
Teleflex Medical  
IDA Business and Technology Park  
Dublin Road, Athlone  
Co. Westmeath, Ireland  
CE 2797  
Arrow International LLC  
Subsidiary of Teleflex Incorporated  
3015 Carrington Mill Blvd.  
Morrisville, NC 27560 USA  
**Teleflex®**  
[www.teleflex.com](http://www.teleflex.com)  
LBL060781 R01 (2024-02)



# MAC™ Two-Lumen Central Venous Access Set with Integral Hemostasis Valve for use with 7 - 7.5 Fr. Catheters



REF

SI-11242

**9** Fr. distal lumen    **2** Lumen    **11.5** cm device length    **.035** inch dia. spring-wire guide