



REF PR-45541-BAS

4.5 Fr. 1 Lumen 55 cm catheter length .018 inch dia. spring-wire guide

Pressure Injectable Arrowgard Blue Advance® One-Lumen PICC

Contents:

- 1: One Lumen TaperFree® Catheter with Arrowgard Blue Advance® Antimicrobial/Antithrombotic Protection¹: 4.5 Fr. (1.58 mm OD) x 55 cm, Pressure Injectable, T-Port Connector, Blue FlexTip® and Placement Wire
- 1: GlideThru™ Peel-Away Sheath: 4.5 Fr. x 2-3/4" (7 cm) Radiopaque over 4.5 Fr. Dilator
- 1: Spring-Wire Guide, Nitinol, Marked: .018" (0.46 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End - Straight Stiff Tip on Other)
- 1: Safety Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW
- 1: Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW
- 1: Syringe: 10 mL Luer-Lock
- 1: Dust Cap: Non-Vented
- 1: SecondSite™ Adjustable Hub: Catheter Clamp
- 1: SecondSite™ Adjustable Hub: Fastener
- 1: SharpsAway® Disposal Cup
- 1: SharpsAway® II Locking Disposal Cup
- 1: Catheter Trimmer
- 1: Safety Scalpel: #11
- 1: Patient ID Card
- 1: Patient Information Booklet
- 2: Paper Tape Measure
- 1: Tourniquet
- 1: Dressing: STATLOCK®² Catheter Stabilization Device

²A registered trademark of C. R. Bard, Inc.

All components are CE 2797 unless otherwise noted.

¹Licensed under US Patent No. 7,329,412.

Rx only

Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. www.teleflex.com/IFU

California Prop. 65

⚠ **WARNING:** Cancer and Reproductive Harm.

www.P65Warnings.ca.gov

Not made with natural rubber latex.

Store below 25°C (77°F). Avoid excessive heat above 30°C (86°F).

Fluid path components are non-pyrogenic.

Contraindications: The Pressure Injectable Arrowgard Blue Advance antimicrobial/antithrombotic catheter is contraindicated:

- for patients with known hypersensitivity to chlorhexidine
- in the presence of device related infection in the intended insertion vessel or catheter pathway
- in the presence of thrombosis in the intended insertion vessel or catheter pathway



Lumen

Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	Pump Flow Rate‡ (mL/hr)	MAX Pressure Injection Flow Rate** (mL/sec)
Distal (17 Ga.)	0.59	913	5130	5

* Priming volumes are approximate and are done without accessories.

† Flow rate values are approximate and are determined using deionized water at 100 cm head height.

‡ Pump flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.

** Pressure injection flow rates are determined at the injector pressure setting of 300 psi maximum using media of 11.8 centipoise viscosity, with 152 cm pressure tubing.



LOT



EU Authorized Representative and Importer:

EC REP Teleflex Medical
IDA Business and
Technology Park
Dublin Road, Athlone
Co. Westmeath, Ireland
2797

Arrow International LLC
Subsidiary of Teleflex Incorporated
3015 Carrington Mill Blvd.
Morrisville, NC 27560 USA

Teleflex®
www.teleflex.com