

Pressure Injectable Arrowg+ard Blue Plus[®] Four-Lumen CVC

Contents:

- 1: Arrowg+ard Blue Plus^{®1} Four-Lumen Catheter: 8.5 Fr. (3.0 mm OD) x 16 cm, Pressure Injectable
- Spring-Wire Guide, Nitinol, Marked: .032" (0.81 mm) dia. x 23-5/8" (60 cm) (Straight Soft Tip on One End - "J" Tip on Other) with Arrow® GlideWheel™ Wire Advancer
- Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopague over 20 Ga. RW Introducer 1: Needle
- Injection Needle: SafetyGlide^{™2} 25 Ga. x 1" (2.54 cm) 1:
- Blunt Fill Needle: 18 Ga. x 1-1/2" (3.81 cm) 1:
- 1: Pressure Transduction Probe
- Introducer Needle: Echogenic 18 Ga. x 2-1/2" (6.35 cm) XTW and 5 mL Arrow® 1: Raulerson Spring-Wire Introduction Syringe
- Injection Needle: SafetyGlide^{™2} 22 Ga. x 1-1/2" (3.81 cm) and 5 mL Luer-Slip 1: Svringe
- Syringe: 3 mL Luer-Lock
- Tissue Dilator: 10 Fr. (3.5 mm) x 10.2 cm 1:
- 4. 10 mL Luer-Lock Pre-Filled Saline Syringe
- 5 mL 1% Lidocaine HCI Solution and Alcohol Prep 1:
- 3 mL Applicator 2% CHG and 70% IPA ChloraPrep®3 One-Step Solution with 2: Hi-Lite Orange™ Tint
- SecondSite[™] Adjustable Hub: Fastener
- SecondSite™ Adjustable Hub: Catheter Clamp 1:
- SharpsAway® Disposal Cup 1.
- 1: SharpsAway® II Locking Disposal Cup
- Maximal Barrier Drape[™] with 4" fenestration 1.
- Towel 1:
- 1: Needle Holder

±Precaution: Contains Phthalate: DEHP.

Warning: The results of certain animal experiments have shown phthalates to be potentially toxic to reproduction. Proceeding from the present state of scientific knowledge, risks for male premature infants cannot be excluded in the case of long-term exposure or application. Medical products containing phthalates should be used only temporarily with pregnant women, nursing mothers, babies and infants.

Not made with natural rubber latex. Store between 20 - 25°C (68 - 77°F)

Arrow International provides the enclosed medication label(s) for your convenience. Please ensure that these labels are applied to the correct syringe and corresponding medication

Contraindications: The Arrowg+ard Blue Plus antimicrobial catheter is contraindicated for patients with known hypersensitivity to chlorhexidine, silver sulfadiazine and/or sulfa drugs.



- 1: Safety Scalpel: #11 1: Checklist/CLIP Sheet
- 1: Sterile Procedure Sign
- Medication Label: 1% Lidocaine 1:
- 2: Gauze Pad: 2" x 2" (5 cm x 5 cm)
- Gauze Pad: 4" x 4" (10 cm x 10 cm) 5:
- Surgical Apparel: Impervious Gown 1.
- Dressing: I.V. Advanced 8.5 cm x 11.5 cm 1:
- Dressing: BIOPATCH®4 1:
- 1: Transducer Cover: 14 cm x 147 cm
- Surgical Apparel: Mask with Eve Shield 1:
- Surgical Apparel: Bouffant Cap 1:
- Suture: 3-0 Silk with Curved Needle 1:
- Tubing±: 20" (50 cm) Extension 1:
- 1: HemoHopper® Fluid Receptacle
- ¹Licensed under US Patent No. 7.329.412.
- ²A trademark of Becton, Dickinson and Company
- 3A registered trademark of CareFusion or one of its subsidiaries.
- ⁴A registered trademark of Johnson & Johnson Corporation.
- Gravity MAX Pressure Primina Flow Injection Volume Ratet Flow Rate** (mL) (mL/hr) (mL/sec) Lumen Distal (16 Ga.) 0.38 2781 5 Medial 1 (14 Ga.) Medial 2 (18 Ga.) 4980 10 0.45 0.35 1453 not rated 0.38 1619 not rated Proximal (18 Ga.)

* Priming volumes are approximate and are done without accessories.

+ Flow rate values are approximate and are determined using deionized water at 100 cm head height.

** Pressure injection flow rates are determined using media of 11.8 centipoise (cP) viscosity, with 152 cm administration set tubing.

LBL061289 R00 (2022-01)

Erg[®]Pack. Complete