

# ARROW



REF PS-01671

## One-Lumen PICC

### Contents:

- 1: One Lumen Catheter: 16 Ga. (1.80 mm OD) x 70 cm, Side Holes, Blue FlexTip®, Contamination Guard
- 1: Peel-Away Catheter: 14 Ga. x 2-1/2" (6.35 cm) Radiopaque over 15 Ga. TW Introducer Needle and 3 mL Luer-Slip Syringe
- 1: Dust Cap: Non-Vented
- 1: SecondSite™ Adjustable Hub: Fastener
- 1: SecondSite™ Adjustable Hub: Catheter Clamp
- 1: Drape: 18" x 26" (46 cm x 66 cm) with 3" (7.6 cm) fenestration
- 1: Patient Information Booklet
- 1: Patient ID Card
- 1: Dressing: STATLOCK®<sup>1</sup> PICC Plus Catheter Stabilization Device

<sup>1</sup>A registered trademark of C. R. Bard, Inc.

All components are CE 2797 unless otherwise noted.

**Rx only Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. [www.teleflex.com/IFU](http://www.teleflex.com/IFU) California Prop. 65**  
**⚠ WARNING: Cancer and Reproductive Harm.**  
[www.P65Warnings.ca.gov](http://www.P65Warnings.ca.gov)

**Not made with natural rubber latex.** Fluid path components are non-pyrogenic.

**Contraindications: The PICC is contraindicated wherever there is presence of device related infections or presence of thrombosis in the intended insertion vessel or catheter pathway. Clinical assessment of the patient must be completed to ensure no contraindications exist.**

|       | Priming Volume* (mL) | Gravity Flow Rate† (mL/hr) | Pump Flow Rate†† (mL/hr) |
|-------|----------------------|----------------------------|--------------------------|
| Lumen | 0.75                 | 1087                       | 6135                     |

\* Priming volumes are approximate and are done without accessories.  
 † Flow rate values are approximate and are determined using deionized water at 100 cm head height.  
 †† Pump flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.

en sl es



~~LATEX~~ CAS # 7440-48-4

Arrow International LLC  
 Subsidiary of Teleflex Incorporated  
 3015 Carrington Mill Blvd.  
 Morrisville, NC 27560 USA

**Teleflex®**  
[www.teleflex.com](http://www.teleflex.com)

EU Authorized Representative and Importer:

EC REP Teleflex Medical  
 IDA Business and Technology Park  
 Dublin Road, Athlone  
 2797 Co. Westmeath, Ireland

LOT



LBL063139 R01 (2023-09)

# ARROW



REF PS-01671

## One-Lumen PICC

### Contents:

- 1: One Lumen Catheter: 16 Ga. (1.80 mm OD) x 70 cm, Side Holes, Blue FlexTip®, Contamination Guard
- 1: Peel-Away Catheter: 14 Ga. x 2-1/2" (6.35 cm) Radiopaque over 15 Ga. TW Introducer Needle and 3 mL Luer-Slip Syringe
- 1: Dust Cap: Non-Vented
- 1: SecondSite™ Adjustable Hub: Fastener
- 1: SecondSite™ Adjustable Hub: Catheter Clamp
- 1: Drape: 18" x 26" (46 cm x 66 cm) with 3" (7.6 cm) fenestration
- 1: Patient Information Booklet
- 1: Patient ID Card
- 1: Dressing: STATLOCK®<sup>1</sup> PICC Plus Catheter Stabilization Device

<sup>1</sup>A registered trademark of C. R. Bard, Inc.

All components are CE 2797 unless otherwise noted.

**Rx only Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. [www.teleflex.com/IFU](http://www.teleflex.com/IFU) California Prop. 65**  
**⚠ WARNING: Cancer and Reproductive Harm.**  
[www.P65Warnings.ca.gov](http://www.P65Warnings.ca.gov)

**Not made with natural rubber latex.** Fluid path components are non-pyrogenic.

**Contraindications: The PICC is contraindicated wherever there is presence of device related infections or presence of thrombosis in the intended insertion vessel or catheter pathway. Clinical assessment of the patient must be completed to ensure no contraindications exist.**

|       | Priming Volume* (mL) | Gravity Flow Rate† (mL/hr) | Pump Flow Rate†† (mL/hr) |
|-------|----------------------|----------------------------|--------------------------|
| Lumen | 0.75                 | 1087                       | 6135                     |

\* Priming volumes are approximate and are done without accessories.  
 † Flow rate values are approximate and are determined using deionized water at 100 cm head height.  
 †† Pump flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.

en sl es



~~LATEX~~ CAS # 7440-48-4

Arrow International LLC  
 Subsidiary of Teleflex Incorporated  
 3015 Carrington Mill Blvd.  
 Morrisville, NC 27560 USA

**Teleflex®**  
[www.teleflex.com](http://www.teleflex.com)

EU Authorized Representative and Importer:

EC REP Teleflex Medical  
 IDA Business and Technology Park  
 Dublin Road, Athlone  
 2797 Co. Westmeath, Ireland

LOT



LBL063139 R01 (2023-09)



16 Ga. 1 Lumen 70 cm catheter length

REF PS-01671  
 One-Lumen PICC



16 Ga. 1 Lumen 70 cm catheter length

REF PS-01671  
 One-Lumen PICC