

Product of xxx
Packaged in xxx

ARROW



REF CA-11142 (IPNXXXXXX)

9 Fr.
distal
lumen

2 Lumen

10 cm
device
length.035 inch dia.
spring-wire
guide

MAC™ Two-Lumen Central Venous Access Kit with Integral Hemostasis Valve for use with 7 - 8 Fr. Catheters

Contents:

- 1: Two-Lumen Central Venous Access Device: 4" (10 cm) Radiopaque Polyurethane with 9 Fr. Distal Lumen, Integral Hemostasis Valve, Extension Line Clamps
- 1: Spring-Wire Guide, Marked: .035" (0.89 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End - "J" Tip on Other) with Arrow Advancer
- 1: Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer Needle
- 1: Pressure Transduction Probe
- 1: Injection Needle: 25 Ga. x 1" (2.54 cm) and 3 mL Luer-Lock Syringe
- 1: Introducer Needle: 18 Ga. x 2-1/2" (6.35 cm) XTW and 5 mL Arrow® Raulerson Spring-Wire Introduction Syringe
- 1: Injection Needle: 22 Ga. x 1-1/2" (3.81 cm) and 5 mL Luer-Slip Syringe
- 1: Tissue Dilator: 9 Fr. (3.0 mm) x 20.7 cm
- 1: Obturator: 8 Fr.
- 1: Cath-Gard® Catheter Contamination Shield: 80 cm with Tuohy-Borst Adapter, Sterile Tape
- 3: Dust Cap: Non-Vented
- 1: SharpsAway® Disposal Cup
- 1: Drape: 24" x 36" (61 cm x 91 cm) with 4" (10 cm) fenestration
- 1: Drape: 24" x 36" (61 cm x 91 cm)
- 1: Scalpel: #11
- 1: Stopcock: 4-way High-Flow
- 2: Gauze Pad: 2" x 2" (5 cm x 5 cm)
- 5: Gauze Pad: 4" x 4" (10 cm x 10 cm)
- 1: Suture: 3-0 Braided Silk with Straight Needle

All components are CE 2797 unless otherwise noted.

Rx only

Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. www.teleflex.com/IFU

California Prop. 65

⚠ **WARNING:** Cancer and Reproductive Harm.
www.P65Warnings.ca.gov

Not made with natural rubber latex.

Fluid path components are non-pyrogenic.



Lumen

Priming
Volume*
(mL)Gravity
Flow
Rate†
(mL/hr)

Distal (9 Fr.) no catheter	1.65	31440
Distal (9 Fr.) with 8 Fr. catheter	0.9	9040
Proximal (12 Ga.) no catheter	0.6	8970

* Priming volumes are approximate and are done without accessories.

† Flow rate values are approximate and are determined using deionized water at 100 cm head height.

en

nl

fr


de

it

pt

es

sv



LOT Lot Number

YYYY-MM-DD

YYYY-MM-DD



(01)XXXXXXXXXXXXXX
(17)YYMMDD
(11)YYMMDD
(10)Lot Number

EU Authorized
Representative and
Importer:

Teleflex Medical
IDA Business and
Technology Park
Dublin Road, Athlone
Co. Westmeath, Ireland



Arrow International LLC
Subsidiary of Teleflex Incorporated
3015 Carrington Mill Blvd.
Morrisville, NC 27560 USA

Teleflex®
www.teleflex.com

LBL064043 R01 (2024-03)



REF CA-11142

9 Fr.
distal
lumen

2 Lumen

10 cm
device
length.035 inch dia.
spring-wire
guide

MAC™ Two-Lumen Central Venous Access Kit with Integral Hemostasis Valve for use with 7 - 8 Fr. Catheters

ARROW