



REF AK-04550-S

20 Ga. 1 Lumen 12 cm catheter length .025 inch dia. spring-wire guide

Vessel Catheterization

Contents:

- 1: One-Lumen Catheter: 20 Ga. (1.2 mm OD) x 12 cm
- 1: Spring-Wire Guide: .025" (0.64 mm) dia. x 13-3/4" (35 cm) (Straight Soft Tip on One End - "J" Tip on Other)
- 1: Catheter: 20 Ga. x 1-3/4" (4.45 cm) Radiopaque over 22 Ga. TW Introducer Needle
- 1: Injection Needle: 25 Ga. x 1" (2.54 cm)
- 1: Injection Needle: 18 Ga. x 1-1/2" (3.81 cm)
- 1: Introducer Needle: 20 Ga. x 2" (5.08 cm) XTW
- 1: Blunt Fill Needle: 18 Ga. x 1-1/2" (3.81 cm)
- 1: Syringe: 3 mL Luer-Lock
- 1: Syringe: 5 mL Luer-Slip
- 1: Tissue Dilator: 4 Fr. (1.35 mm) x 7.6 cm
- 1: Tissue Dilator: 4 Fr. (1.35 mm) x 4.1 cm
- 1: 5 mL 1% Lidocaine HCl Solution and Alcohol Prep
- 1: 3 mL Applicator 2% CHG and 70% IPA ChloraPrep®<sup>†</sup> One-Step Solution
- 1: Drape: 18" x 26" (46 cm x 66 cm) with 3" (7.6 cm) fenestration
- 1: Scalpel: #11
- 3: Gauze Pad: 2" x 2" (5 cm x 5 cm)
- 1: Suture: 3-0 Braided Silk with Straight Needle

<sup>†</sup>A registered trademark of CareFusion or one of its subsidiaries.

Rx only

Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. [www.teleflex.com/IFU](http://www.teleflex.com/IFU)

California Prop. 65

**WARNING:** Cancer and Reproductive Harm. [www.P65Warnings.ca.gov](http://www.P65Warnings.ca.gov)

Not made with natural rubber latex.

Store between 20 – 25°C (68 – 77°F).

Sterilized using ethylene oxide unless otherwise indicated in the contents list.

Fluid path components are non-pyrogenic.



Lumen

Priming Volume\*

(mL)

Gravity Flow Rate†

(mL/hr)

Lumen	Priming Volume*	Gravity Flow Rate†
Distal (20 Ga.)	0.1	1531

\* Priming volumes are approximate and are done without accessories.

† Flow rate values are approximate and are determined using deionized water at 100 cm head height.

Due To A Temporary Supply Shortage  
A Clear ChloraPrep Has Been Added To This Kit To Replace  
The Orange Tinted ChloraPrep.



LOT



Arrow International LLC  
Subsidiary of Teleflex Incorporated  
3015 Carrington Mill Blvd.  
Morrisville, NC 27560 USA

Teleflex®  
[www.teleflex.com](http://www.teleflex.com)

LBL067087 R00 (2022-07)



REF AK-04550-S

20 Ga. 1 Lumen 12 cm catheter length .025 inch dia. spring-wire guide

Vessel Catheterization