



REF IDLX-45541-CURV

(IPNXXXXXX)

4.5 Fr. 1 Lumen 55 cm catheter length .018 inch dia. spring-wire guide


Pressure Injectable Arrowg+ard Blue Advance® One-Lumen PICC pre-loaded with Arrow® NaviCurve™ Stylet



Contents:

- 1: One Lumen TaperFree® Catheter,††† with Arrowg+ard Blue Advance® Antimicrobial/Antithrombogenic Protection†: 4.5 Fr. (1.58 mm OD) x 55 cm, Pressure Injectable, Arrow® NaviCurve™ Stylet and T-Port Connector
- 1: GlideThru™ Peel-Away Sheath: 4.5 Fr. x 2-3/4" (7 cm) Radiopaque over 4.5 Fr. Dilator
- 1: Spring-Wire Guide, Nitinol, Marked: .018" (0.46 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End - Straight Stiff Tip on Other) with Arrow Advancer
- 1: Introducer Needle: 21 Ga. x 1-1/2" (3.81 cm) TW
- 1: Injection Needle: Eclipse™2 25 Ga. x 1" (2.54 cm)
- 1: Safety Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW
- 1: Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW
- 1: Filter Needle: 18 Ga. TW x 1-1/2" (3.81 cm) with Blunt Fill Tip
- 1: Syringe: 10 mL Luer-Lock
- 1: Syringe: 3 mL Luer-Lock
- 1: Dust Cap: Non-Vented
- 1: SecondSite™ Adjustable Hub: Catheter Clamp
- 1: SecondSite™ Adjustable Hub: Fastener
- 1: SharpsAway® Disposal Cup
- 1: SharpsAway® II Locking Disposal Cup
- 1: Catheter Trimmer
- 1: Towel: 24" x 36" (61 cm x 91 cm)
- 1: Maximal Barrier Drape™ with two 13 cm (5") x 15 cm (6") fenestrations
- 1: Towel
- 1: Safety Scalpel: #11
- 1: Sterile Procedure Sign
- 1: Patient ID Card
- 1: Patient Information Booklet
- 2: Paper Tape Measure
- 1: Tourniquet
- 1: Foam Electrodes (3 per pack)
- 5: Gauze Pad: 2" x 2" (5 cm x 5 cm)
- 10: Gauze Pad: 4" x 4" (10 cm x 10 cm)
- 1: Surgical Apparel: Impervious Gown
- 1: Dressing: I.V. Advanced 8.5 cm x 11.5 cm
- 1: Transducer Cover with Gel and Rubber Bands
- 1: Remote Cover
- 1: Surgical Apparel: Mask with Eye Shield
- 1: Dressing: STATLOCK®3 PICC Plus Catheter Stabilization Device
- 1: Surgical Apparel: Bouffant Cap
- 1: Surgical Apparel: Mask
- 1: Tape: Steri-Strip®4
- 1: Valve: MicroCLAVE™5 Neutral Displacement Connector
- 1: T-Piece Cover

†Licensed under US Patent No. 7,329,412 and 11,219,706.
 ‡A trademark of Becton, Dickinson and Company.
 †A registered trademark of C. R. Bard, Inc.
 ‡A registered trademark of 3M Company.
 †A trademark of ICU Medical, Inc.

 Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	Pump Flow Rate†† (mL/hr)	MAX Pressure Injection Flow Rate** (mL/sec)
Pink (17 Ga.)	0.66	1090	5380	5

* Priming volumes are approximate and are done without accessories.
 † Flow rate values are approximate and are determined using deionized water at 100 cm head height.
 †† Pump flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.
 ** Pressure injection flow rates are determined at the injector pressure setting of 300 psi maximum using media of 11.8 centipoise viscosity, with 152 cm pressure tubing.

