



REF CDA-45563-HPK1A (IPNXXXXXX)

6 Fr. 3 Lumen 55 cm catheter length .018 inch dia. spring-wire guide



Pressure Injectable Arrowg+ard Blue Advance® Three-Lumen PICC

Contents:

- 1: Three Lumen TaperFree® Catheter with Arrowg+ard Blue Advance® Antimicrobial/Antithrombogenic Protection¹: 6 Fr. (2.11 mm OD) x 55 cm, Pressure Injectable, T-Port Connector, Blue FlexTip® and Placement Wire

1: GlideThru™ Peel-Away Sheath: 6 Fr. x 2-3/4" (7 cm) Radiopaque over 6 Fr. Dilator

1: Spring-Wire Guide, Nitinol, Marked: .018" (0.46 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End - Straight Stiff Tip on Other) with Arrow Advancer

1: Injection Needle: Eclipse™² 25 Ga. x 1" (2.54 cm)

1: Safety Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW

1: Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW

1: Syringe: 10 mL Luer-Lock

1: Syringe: 3 mL Luer-Lock

3: Dust Cap: Non-Vented

1: SecondSite™ Adjustable Hub: Fastener

1: SecondSite™ Adjustable Hub: Catheter Clamp

1: SharpsAway® II Locking Disposal Cup

1: SharpsAway® Disposal Cup

1: Catheter Trimmer

1: Towel: 24" x 36" (61 cm x 91 cm)

1: Maximal Barrier Drape™ with two 13 cm (5") x 15 cm (6") fenestrations

1: Towel

1: Safety Scalpel: #11

1: Patient ID Card

1: Patient Information Booklet

2: Paper Tape Measure

1: Tourniquet

10: Gauze Pad: 2" x 2" (5 cm x 5 cm)

10: Gauze Pad: 4" x 4" (10 cm x 10 cm)

1: Dressing: Tegaderm®³ 15.5 cm x 10 cm

1: Surgical Apparel: Impervious Gown
- 1: Dressing: STATLOCK®⁴ Catheter Stabilization Device

1: Surgical Apparel: Mask with Eye Shield

1: Surgical Apparel: Bouffant Cap

1: Surgical Apparel: Mask

1: Tape: Steri-Strip®³

1: HemoHopper® Fluid Receptacle
- ¹Licensed under US Patent No. 7,329,412.


²A trademark of Becton, Dickinson and Company.

³A registered trademark of 3M Company.

⁴A registered trademark of C. R. Bard, Inc.
- All components are CE 2797 unless otherwise noted.



Not made with natural rubber latex.
Store below 25°C (77°F). Avoid excessive heat above 30°C (86°F).
Contraindications: The Pressure Injectable Arrowg+ard Blue Advance antimicrobial/antithrombogenic catheter is contraindicated:
• for patients with known hypersensitivity to chlorhexidine
• in the presence of device related infection in the intended insertion vessel or catheter pathway
• in the presence of thrombosis in the intended insertion vessel or catheter pathway

 Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	Pump Flow Rate†† (mL/hr)	MAX Pressure Injection Flow Rate** (mL/sec)
Distal (18 Ga.)	0.61	880	5360	6
Medial (19 Ga.)	0.45	244	2090	not rated
Proximal (19 Ga.)	0.47	257	2160	not rated

* Priming volumes are approximate and are done without accessories.
† Flow rate values are approximate and are determined using deionized water at 100 cm head height.
†† Pump flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.
** Pressure injection flow rates are determined at the injector pressure setting of 300 psi maximum using media of 11.8 centipoise viscosity, with 152 cm pressure tubing.

