

Dimensions:  
Width: 6-9/64" (15.6 cm)  
Length: 8-21/32" (22 cm)

PB-50014-100B  
PBZ-50014-103A

T50014100A6 Dimensions: 5 7/8" x 7 5/8"

Product of xxx  
Packaged in xxx

ARROW®



REF CS-04306-EJ

(IPNXXXXXX)


# One-Lumen CVC

## Contents:

- 1: One-Lumen Catheter: 16 Ga. (1.8 mm OD) x 16 cm
- 1: Spring-Wire Guide, Marked: .032" (0.81 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End - "J" Tip on Other) with Arrow Advancer
- 1: Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer Needle
- 1: Introducer Needle: 18 Ga. x 2-1/2" (6.35 cm) XTW
- 1: Pressure Transduction Probe
- 1: Arrow® Raulerson Spring-Wire Introduction Syringe: 5 mL
- 1: Tissue Dilator: 7 Fr. (2.3 mm) x 10.2 cm
- 1: Dust Cap: Non-Vented
- 1: SecondSite™ Adjustable Hub: Fastener
- 1: SecondSite™ Adjustable Hub: Catheter Clamp

All components are CE 2797 unless otherwise noted.

**Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. [www.teleflex.com](http://www.teleflex.com)/IFU Not made with natural rubber latex. Fluid path components are non-pyrogenic.**



Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)
Distal (16 Ga.)	0.4	3013

\* Priming volumes are approximate and are done without accessories.  
† Flow rate values are approximate and are determined using deionized water at 100 cm head height.

CAS # 7440-48-4

EU Authorized Representative and Importer:

EC REP  
CE 2797  
Teleflex Medical  
IDA Business and  
Technology Park  
Dublin Road, Athlone  
Co. Westmeath, Ireland

Arrow International LLC  
Subsidiary of Teleflex Incorporated  
3015 Carrington Mill Blvd.  
Morrisville, NC 27560 USA

Teleflex®  
[www.teleflex.com](http://www.teleflex.com)

LBL074417 R00 (2024-02)

LOT LotNumber

YYYY-MM-DD

YYYY-MM-DD



(01)XXXXXXXXXXXXXX  
(17)YYMMDD  
(11)YYMMDD  
(10)LotNumber



16 Ga. 1 Lumen

16 cm catheter length

.032 inch dia. spring-wire guide

REF CS-04306-EJ

One-Lumen CVC

ARROW®