

**REF IDLX-45552-CURV** (IPNXXXXXX)

> 2 Lumen

## Pressure Injectable Arrowg+ard Blue Advance® Two-Lumen PICC pre-loaded with Arrow<sup>®</sup> NaviCurve<sup>™</sup> Stylet

Contents:

- 1: Two Lumen TaperFree® Catheter<sup>‡‡‡</sup> with Arrowg+ard Blue Advance® Antimicrobial/Antithrombogenic Protection1: 5.5 Fr. (1.88 mm OD) x 55 cm, Pressure Injectable, Arrow® NaviCurve™ Stylet and T-Port Connector GlideThru™ Peel-Away Sheath: 5.5 Fr. x 2-3/4" (7 cm) Radiopaque over 5.5 1. Fr. Dilator Spring-Wire Guide, Nitinol, Marked: .018" (0.46 mm) dia. x 17-3/4" (45 cm)
- (Straight Soft Tip on One End Straight Stiff Tip on Other) with Arrow Advancer Introducer Needle: 21 Ga. x 1-1/2" (3.81 cm) TW
- Injection Needle: Eclipse<sup>™2</sup> 25 Ga. x 1" (2.54 cm)
- Safety Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW
- Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW
- Filter Needle: 18 Ga. TW x 1-1/2" (3.81 cm) with Blunt Fill Tip
- Svringe: 10 mL Luer-Lock 1.
- Syringe: 3 mL Luer-Lock
- 2: Dust Cap: Non-Vented
- SecondSite™ Adjustable Hub: Fastener 1.
- SecondSite<sup>™</sup> Adjustable Hub: Catheter Clamp
- SharpsAway® Disposal Cup
- SharpsAway® II Locking Disposal Cup Catheter Trimmer 1.
- Towel: 24" x 36" (61 cm x 91 cm)
- Maximal Barrier Drape™ with two 13 cm (5") x 15 cm (6") fenestrations 1.
- Towel 1.
- Safety Scalpel: #11
- 1. Patient ID Card
- Sterile Procedure Sign
- 1: Patient Information Booklet

## +++MR Safe

Not made with natural rubber latex.

Store below 25°C (77°F). Avoid excessive heat above 30°C (86°F)

- Contraindications: The Pressure Injectable Arrowg+ard Blue Advance antimicrobial/antithrombogenic catheter is contraindicated:
- for patients with known hypersensitivity to chlorhexidine
- in the presence of device related infection in the intended insertion vessel or catheter pathway
- in the presence of thrombosis in the intended insertion vessel or catheter pathway

2: Paper Tape Measure

55 cm catheter length

- 1: Tourniquet
- 1: Foam Electrodes (3 per pack)
- 5: Gauze Pad: 2" x 2" (5 cm x 5 cm) 10: Gauze Pad: 4" x 4" (10 cm x 10 cm)
- Surgical Apparel: Impervious Gown 1:
- Dressing: I.V. Advanced 8.5 cm x 11.5 cm 1.
- Transducer Cover with Gel and Rubber Bands 1:
- Remote Cover
- 1. 1: Surgical Apparel: Mask with Eye Shield
- Dressing: STATLOCK®3 PICC Plus Catheter Stabilization Device 1:

.018 inch dia. spring-wire guide

- Surgical Apparel: Bouffant Cap 1:
- Surgical Apparel: Mask 1:
- Tape: Steri-Strip®4 1.
- 2: Valve: MicroCLAVE<sup>™5</sup> Neutral Displacement Connector
- 1. T-Piece Cover
- <sup>1</sup>Licensed under US Patent No. 7,329,412 and 11,219,706.
- <sup>2</sup>A trademark of Becton, Dickinson and Company
- <sup>3</sup>A registered trademark of C. R. Bard, Inc.
- <sup>4</sup>A registered trademark of 3M Company.
- <sup>5</sup>A trademark of ICU Medical. Inc.
- All components are CE 2797 unless otherwise noted.

<b>O</b> Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	Pump Flow Rate†† (mL/hr)	MAX Pressure Injection Flow Rate** (mL/sec)
Pink (18 Ga.)	0.56	350	3050	5
White (18 Ga.)	0.58	360	3100	5

\* Priming volumes are approximate and are done without accessories.

† Flow rate values are approximate and are determined using deionized water at 100 cm head height.

++ Pump flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.

\*\* Pressure injection flow rates are determined at the injector pressure setting of 300 psi maximum using media of 11.8 centipoise viscosity, with 152 cm pressure tubing.

(////CAS # 7440-48-4 \*\*\* MR LBL078817 R00 (2025-03)

