

## Pressure Injectable Arrowg+ard Blue Advance<sup>®</sup> One-Lumen PICC pre-loaded with Arrow<sup>®</sup> NaviCurve<sup>™</sup> Stylet

Contents:

- 1: One Lumen TaperFree® Catheter<sup>‡‡‡</sup> with Arrowg+ard Blue Advance® 2: Paper Tape Measure Antimicrobial/Antithrombogenic Protection1: 4.5 Fr. (1.58 mm OD) x 55 cm, 1: Tourniquet Pressure Injectable, Arrow® NaviCurve™ Stylet and T-Port Connector 1: Foam Electrodes (3 per pack) GlideThru™ Peel-Away Sheath: 4.5 Fr. x 2-3/4" (7 cm) Radiopaque over 4.5 5: Gauze Pad: 2" x 2" (5 cm x 5 cm) 1. Fr. Dilator 10: Gauze Pad: 4" x 4" (10 cm x 10 cm) Spring-Wire Guide, Nitinol, Marked: .018" (0.46 mm) dia. x 17-3/4" (45 cm) Surgical Apparel: Impervious Gown 1: (Straight Soft Tip on One End - Straight Stiff Tip on Other) with Arrow Advancer Dressing: I.V. Advanced 8.5 cm x 11.5 cm 1. Introducer Needle: 21 Ga. x 1-1/2" (3.81 cm) TW Transducer Cover with Gel and Rubber Bands 1: Injection Needle: Eclipse<sup>™2</sup> 25 Ga. x 1" (2.54 cm) 1. Remote Cover Safety Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW 1. Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW 1: Filter Needle: 18 Ga. TW x 1-1/2" (3.81 cm) with Blunt Fill Tip Surgical Apparel: Bouffant Cap 1: Svringe: 10 mL Luer-Lock Surgical Apparel: Mask 1: 1. Syringe: 3 mL Luer-Lock Tape: Steri-Strip®4 1. Dust Cap: Non-Vented SecondSite<sup>™</sup> Adjustable Hub: Catheter Clamp 1. T-Piece Cover SecondSite<sup>™</sup> Adjustable Hub: Fastener
- 1: SharpsAway® Disposal Cup
- 1: SharpsAway® II Locking Disposal Cup
- 1: Catheter Trimmer
- : Towel: 24" x 36" (61 cm x 91 cm)
- 1: Maximal Barrier Drape™ with two 13 cm (5") x 15 cm (6") fenestrations
- 1: Towel
- 1: Safety Scalpel: #11
- 1: Patient ID Card
- 1: Sterile Procedure Sign
- 1: Patient Information Booklet

## +++MR Safe

Not made with natural rubber latex.

Store below 25°C (77°F). Avoid excessive heat above 30°C (86°F)

- Contraindications: The Pressure Injectable Arrowg+ard Blue Advance antimicrobial/antithrombogenic catheter is contraindicated:
- for patients with known hypersensitivity to chlorhexidine
- in the presence of device related infection in the intended insertion vessel or catheter pathway
- in the presence of thrombosis in the intended insertion vessel or catheter pathway

- 1: Surgical Apparel: Mask with Eye Shield
  2: Dressing: STATLOCK® PICC Plus Catheter Stabilization Device
  2: Surgical Apparel: Bouffant Cap
  2: Surgical Apparel: Mask
  2: Tape: Steri-Strip®4
  2: Valve: MicroCLAVE TM5 Neutral Displacement Connector
  2: T-Piece Cover
  3: Licensed under US Patent No. 7,329,412 and 11,219,706.
  2: A trademark of Becton, Dickinson and Company.
  3: A produced tendematic of C. P. Bard, Isa
- <sup>3</sup>A registered trademark of C. R. Bard, Inc.
- <sup>4</sup>A registered trademark of 3M Company <sup>5</sup>A trademark of ICU Medical. Inc.
- All components are CE 2797 unless otherwise noted.

C	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	Pump Flow Rate†† (mL/hr)	MAX Pressure Injection Flow Rate** (mL/sec)
Pink (17 Ga.)	0.66	1090	5380	5

\* Priming volumes are approximate and are done without accessories.

† Flow rate values are approximate and are determined using deionized water at 100 cm head height.

†† Pump flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.

\*\* Pressure injection flow rates are determined at the injector pressure setting of 300 psi maximum using media of 11.8 centipoise viscosity, with 152 cm pressure tubing.



