

REF **SAC-01220-B** (IPNXXXXXX)
20 Ga. **12** cm catheter length **.021** inch dia. spring-wire guide

ARROW
ErgoPack®


Arterial Catheterization

Contents:

- | | |
|------------------------------------------------------------------------------------------------------------|-------------------------------------------------------------------------------|
| 1: Catheter: 20 Ga. x 12 cm Radiopaque PEBA | 1: Surgical Apparel: Impervious Gown |
| 1: Spring-Wire Guide, Marked: .021" (0.53 mm) dia. x 13-3/4" (35 cm) (Straight Soft Tip on Both Ends) | 1: Dressing: I.V. Advanced 10 cm x 12 cm |
| 1: Introducer Needle: 20 Ga. x 2-3/4" (7 cm) XTW | 1: Dressing: BIOPATCH® ³ |
| 1: Injection Needle: Eclipse™ ¹ 23 Ga. x 1-1/2" (3.81 cm) | 1: Transducer Cover: 14 cm x 147 cm |
| 1: Injection Needle: Eclipse™ ¹ 25 Ga. x 1" (2.54 cm) | 1: Surgical Apparel: Bouffant Cap |
| 1: Blunt Fill Needle: 18 Ga. x 1-1/2" (3.81 cm) | 1: Surgical Apparel: Mask with Eye Shield |
| 1: Syringe: 3 mL Luer-Lock | 1: Suture: 3-0 Silk with Curved Needle |
| 1: Syringe: 5 mL Luer-Slip | 1: HemoHopper® Fluid Receptacle |
| 1: 5 mL 1% Lidocaine HCl Solution and Alcohol Prep | ¹ A trademark of Becton, Dickinson and Company. |
| 2: 3 mL Applicator 2% CHG and 70% IPA Chloraprep® ² One-Step Solution with Hi-Lite Orange™ Tint | ² A registered trademark of CareFusion or one of its subsidiaries. |
| 1: Dust Cap: Non-Vented | ³ A registered trademark of Johnson & Johnson Corporation. |
| 1: SharpsAway® Disposal Cup | |
| 1: SharpsAway® II Locking Disposal Cup | |
| 1: Maximal Barrier Drape™ with 4" fenestration | |
| 1: Towel | |
| 1: Needle Holder | |
| 1: Safety Scalpel: #11 | |
| 1: Checklist Sheet | |
| 1: Medication Label: 1% Lidocaine | |
| 4: Gauze Pad: 2" x 2" (5 cm x 5 cm) | |
| 4: Gauze Pad: 4" x 4" (10 cm x 10 cm) | |

Catheter Priming Volume: 0.20 mL. Priming volume is approximate.

Not made with natural rubber latex.

Store between 20 – 25°C (68 – 77°F).

Arrow International provides the enclosed medication label(s) for your convenience. Please ensure that these labels are applied to the correct syringe and corresponding medication.

