

Radial Artery Catheterization

RA CATH KIT: 20 GA X 1 1/2 IN (3.81 CM)

SKU / Article #: ASK-04120-HF

Highlights

Sharps Safety

Considerations

Not made with natural rubber latex.
Includes Lidocaine

Sales Quantity/Case: 5

Representative photo(s): actual product specifications may vary from that shown in the photo

Radial Artery Catheterization
SKU / Article #: ASK-04120-HF





Quantity	Component Description
1	Indwelling Catheter: 20 Ga. x 3.81 cm Radiopaque Polyurethane with Integral Suture Wing
1	Spring-Wire Guide: .025" (0.64 mm) dia. x 13-1/8" (33 cm) (Straight Soft Tip on Both Ends)
1	Injection Needle: SafetyGlide ^{TM1} 18 Ga. x 1-1/2" (3.81 cm)
1	Introducer Needle: 20 Ga. x 1-1/2" (3.81 cm) XTW
1	Injection Needle: SafetyGlide ^{TM1} 25 Ga. x 1" (2.54 cm)
1	Blunt Fill Needle: 18 Ga. x 1-1/2" (3.81 cm)
1	Syringe: 3 mL Luer-Lock
1	5 mL 1% Lidocaine HCl Solution and Alcohol Prep
1	3 mL Applicator 2% CHG and 70% IPA ChloraPrep ^{®2} One-Step Solution
1	3 mL Applicator 2% CHG and 70% IPA ChloraPrep ^{®2} One-Step Solution with Hi-Lite Orange TM Tint
1	SharpsAway [®] Disposal Cup
1	SharpsAway [®] II Locking Disposal Cup
1	Drape: 41" x 55" with 4" fenestration
1	Needle Holder
1	Forceps
1	Safety Scalpel #11
4	Gauze Pad: 2" x 2" (5 cm x 5 cm)
3	Gauze Pad: 4" x 4" (10 cm x 10 cm)
1	Dressing: Tegaderm ^{®3} I.V. 11.5 cm x 8.9 cm
1	Surgical Apparel: Impervious Gown
1	Surgical Apparel: Bouffant Cap
1	Surgical Apparel: Mask with Eye Shield
1	Suture: 3-0 Silk with Curved Needle

¹ A trademark of Becton, Dickinson and Company.

² A registered trademark of CareFusion or one of its subsidiaries.

³ A registered trademark of 3M Company.

Documents



The products in this catalog may not be available in all countries. Please contact your local representative.

Indications, contraindications, warnings and instructions for use can be found in the Product Instructions for Use available under the Documents section on this page, where an IFU is available.