



Pressure Injectable Three-Lumen CVC Kit

MULTI LUMEN PI CVC KIT: 3L 7 FR X 6IN

SKU / Article #: ASK-12703-SPH1

Highlights

Sharps Safety

Considerations

Not made with natural rubber latex.

Sales Quantity/Case: 5

Representative photo(s): actual product specifications may vary from that shown in the photo

Pressure Injectable Three-Lumen CVC Kit	
SKU / Article #: ASK-12703-SPH1	
Quantity	Component Description
1	Three-Lumen Catheter: 7 Fr. (2.5 mm OD) x 16 cm, Pressure Injectable
1	Spring-Wire Guide, Marked: .032" (0.81 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End - "J" Tip on Other) with Arrow Advancer
1	Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer Needle
1	Pressure Transduction Probe
1	Introducer Needle: Echogenic 18 Ga. x 2-1/2" (6.35 cm) XTW and 5 mL Arrow® Raulerson Spring-Wire Introduction Syringe
1	Injection Needle: Eclipse™ ¹ 23 Ga. x 1-1/2" (3.81 cm) and 5 mL Luer-Slip Syringe
1	Injection Needle: Eclipse™ ¹ 25 Ga. x 1" (2.54 cm) and 3 mL Luer-Lock Syringe
1	Syringe: 10 mL Luer-Lock
1	Tissue Dilator: 8.5 Fr. (2.8 mm) x 10.2 cm
1	SecondSite™ Adjustable Hub: Fastener
1	SecondSite™ Adjustable Hub: Catheter Clamp
1	SharpsAway® II Locking Disposal Cup

Pressure Injectable Three-Lumen CVC Kit
SKU / Article #: ASK-12703-SPH1

Quantity	Component Description
1	SharpsAway® Disposal Cup
1	Maximal Barrier Drape™ with 4" fenestration
1	Towel
1	Needle Holder
1	Safety Scalpel: #11
1	Checklist/CLIP Sheet
1	Sterile Procedure Sign
2	Gauze Pad: 3" x 3" (7.6 cm x 7.6 cm)
2	Gauze Pad: 2" x 2" (5 cm x 5 cm)
5	Gauze Pad: 4" x 4" (10 cm x 10 cm)
1	Surgical Apparel: Impervious Gown
1	Dressing: I.V. Advanced 8.5 cm x 11.5 cm
1	Surgical Apparel: Bouffant Cap
1	Surgical Apparel: Mask with Eye Shield
1	Suture: 3-0 Silk with Curved Needle
3	Valve: MaxZero™ ² Needleless Connector
1	HemoHopper® Fluid Receptacle

¹ A trademark of Becton, Dickinson and Company.

² A trademark of CareFusion or one of its subsidiaries.

Documents



The products in this catalog may not be available in all countries. Please contact your local representative.

Indications, contraindications, warnings and instructions for use can be found in the Product Instructions for Use available under the Documents section on this page, where an IFU is available.